



## wwPDB EM Validation Summary Report ⓘ

Jul 12, 2021 – 11:03 am BST

EMDB ID : EMD-12325  
Title : In situ cryo-electron tomogram of the Synechocystis F4E VIPP1 mutant grown in high light  
Authors : Wietrzynski, W.; Klumpe, S.; Heinz, S.; Spaniol, B.; Schaffer, M.; Rast, A.; Nickelsen, J.; Schroda, M.; Engel, B.D.  
Deposited on : 2021-02-10  
Resolution : Not provided

This is a wwPDB EM Validation Summary Report for a publicly released PDB entry.

We welcome your comments at [validation@mail.wwpdb.org](mailto:validation@mail.wwpdb.org)

A user guide is available at

<https://www.wwpdb.org/validation/2017/EMTomogramValidationReportHelp>

with specific help available everywhere you see the ⓘ symbol.

---

The following versions of software and data (see [references ⓘ](#)) were used in the production of this report:

EMDB validation analysis : 0.0.0.dev84  
Validation Pipeline (wwPDB-VP) : 2.22


# 1 Experimental information

Property	Value	Source
EM reconstruction method	TOMOGRAPHY	Depositor
Imposed symmetry	Not Provided	
Number of tilted images used	61	Depositor
Resolution determination method	Not provided	
CTF correction method	Not provided	
Microscope	FEI TITAN KRIOS	Depositor
Voltage (kV)	300	Depositor
Electron dose ( $e^-/\text{\AA}^2$ )	2.0	Depositor
Minimum defocus (nm)	4.0	Depositor
Maximum defocus (nm)	5.5	Depositor
Magnification	42000.0	Depositor
Image detector	GATAN K2 SUMMIT (4k x 4k)	Depositor
Maximum voxel value	2523.000	Depositor
Minimum voxel value	-5742.000	Depositor
Average voxel value	119.042	Depositor
Voxel value standard deviation	205.152	Depositor
Recommended contour level	Not applicable	
Tomogram size (Å)	13066.24, 13066.24, 6533.12	wwPDB
Tomogram dimensions	928, 928, 464	wwPDB
Tomogram angles (°)	90.0, 90.0, 90.0	wwPDB
Grid spacing (Å)	14.08, 14.08, 14.08	Depositor

## 2 Tomogram visualisation [i](#)


This section contains visualisations of the EMDB entry EMD-12325. These allow visual inspection of the internal detail of the tomogram and identification of artifacts.

### 2.1 Orthogonal projections [i](#)



The images above show the tomogram projected in three orthogonal directions.

### 2.2 Central slices [i](#)




The images above show central slices of the tomogram in three orthogonal directions.


## 2.3 Largest variance slices [i](#)



X Index: 807



Y Index: 384



Z Index: 235

The images above show the largest variance slices of the tomogram in three orthogonal directions.


## 2.4 Mask visualisation [i](#)

This section was not generated. No masks/segmentation were deposited.

### 3 Tomogram analysis [i](#)

This section contains the results of statistical analysis of the tomogram.

#### 3.1 Voxel-value distribution [i](#)



The voxel-value distribution is plotted in 128 intervals along the x-axis. The y-axis is logarithmic.